

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Melavilvarayanallur	Details About Robust Analysis
Lab No	12921	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	110.3 mg/dL	-1.27	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.49 mg/dL	1.89	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	115.5 mg/dL	0.67	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மேலவில்லிவராயநல்லூர்
 தி.மலை மாவட்டம்-606906

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&y