

Temperature Unit :			
S. no.	Expected Temperature	Measured Temperature	Remarks
1	37 deg. Celisius	114.9 ohm: (37.25 deg., Celisius)	
	Wait for 30 minutes approximately		
	Used: Calibrated PT 100 Sensor		
Aspiration Volume:			
	Set Vol.	Aspiration	Remarks
	450ul	450ul	Aspiration Vol., are good.
	500ul	495ul	Adjustments not required.
	1000ul	1000ul	
	1500ul	1500ul	
Comments :			
Instrument had been calibrated with calibrated multimeter and PT 100 sensor, whose tracibility certificates are attached with this report. Instrument is in working well within the specification.			
.....			
Service Engineer signature:			