


## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. PHC, Melavilvarayanallur		<b>Details About Robust Analysis</b>							
<b>Lab No</b>	12921		<b>Detail About Monthly Summary</b>							
<b>Month</b>	August		<b>Detail about SDI</b>							
<b>Year</b>	2023									
<b>Constituent Group</b>	Chemistry III									
Click on the analyte to view Graphical Data			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 19/08/2023							
			Date of Report Published : 01/09/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	176.4 mg/dL	0.09	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	117.3 mg/dL	962.83	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	1.6 mg/dL	-24.18	0.22
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
**மருத்துவ அலுவலர்**  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 மேல்வில்வராயநல்லூர்  
 தி.மலை மாவட்டம்-605906.