



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023 Calibration valid : 25.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, V.V.M. HOME,COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNIPIPETTE F3
iii Value	10 µl -100 µl
Serial No	RW01629

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.05	50.08	100.19
2	10.09	50.07	100.11
3	10.04	50.01	100.09
4	10.07	50.06	100.06
5	10.09	49.95	100.08
6	10.08	50.03	100.13

For Heart Beat Medical Systems,
(Authorized signatory) 

