



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date: 26.09.2023 Calibration valid: 25.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, V.V.M. HOME,COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	MICROPET
ii Model	-----
iii Value	100µl-1000µl
iv Serial No	UPHC/VVM.PI/002

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.09	1000.11
2	100.03	500.06	1000.13
3	100.04	500.07	1000.09
4	100.06	500.05	1000.07
5	100.09	499.93	1000.04
6	100.05	500.02	1000.15

For Heart Beat Medical Systems,

*S. Anje*

(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com