



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 20.09.2023 Calibration valid : 19.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, PRIMARY HEALTH CENTRE , ANANTHAL, KALARPALAYAM , THIRUVANNAMALAI-606 902.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINEPITTE – F3
iii Value	100µl-1000µl
iv Serial No	4640040

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.04	500.11	1000.13
2	100.01	500.08	1000.21
3	100.02	500.09	1000.09
4	100.10	500.12	1000.12
5	100.05	500.05	1000.16
6	100.02	500.05	1000.12

For Heart Beat Medical Systems,

  
(Authorized signatory)

