



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date : 20.09.2023
		Calibration valid : 19.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE , ANANTHAL, KALARPALAYAM , THIRUVANNAMALAI-606 902.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINEPITTE F3
iii	Value	10 µl -100 µl
iv	Serial No	4640040-N

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.03	100.14
2	10.04	50.05	100.12
3	10.01	50.02	100.02
4	10.02	50.08	100.08
5	10.00	50.02	100.09
6	10.05	50.07	100.05

For Heart Beat Medical Systems

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com