

Name: Govt. PHC, Sittling [Details About Robust Analysis](#)
 ID: 12107 [Detail About Monthly Summary](#)
 Date: May 2023 [Detail about SDI](#)
 Student Group: Chemistry III
[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)
 Date of Result Entered : 19/05/2023
 Date of Report Published : 05/06/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	180.5 mg/dL	-0.04	0.44
UREA III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110.7 mg/dL	-0.67	0.27

SDI Range	Interpretation
SDI ≥ -1.00 to $+1.00$	Excellent.
SDI $\geq \pm 1.01$ to ± 2.00	Good.
SDI $\geq \pm 2.01$ to ± 2.99	Accept with caution. Warning Signal.
SDI $\geq \pm 3.0$	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out