



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kattabettu

[Details About Robust Analysis](#)

Lab No 12471

[Detail About Monthly Summary](#)

Month March

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Any Analyser (Automation / Semi Automation) | 74 | 92.52 | 7.71 | 7.13 | 93 mg/dL | 0.07 | 1.66 |
| 2 | UREA III | Urease UV / GLDH III | Any Analyser (Automation / Semi Automation) | 12 | 55.01 | 12.13 | 6.67 | 55 mg/dL | 0.00 | 3.85 |
| 3 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Any Analyser (Automation / Semi Automation) | 70 | 1.32 | 7.52 | 0.10 | 1.4 mg/dL | 0.81 | 0.02 |
| 4 | CHOLESTROL III | CHOD-PAP III | Any Analyser (Automation / Semi Automation) | 148 | 111.55 | 4.29 | 4.78 | 110 mg/dL | -0.32 | 0.79 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

| Links | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |

Site Maintained by
 Swiftware Solutions

Best Vi

(Handwritten Signature)
 MLD
 GOVT. PHC, KATTABETTU
 KATTABETTU
 KOTAGIRI, NELLORE