



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Karisalaivanasala  
 Lab No: 7071  
 Month: August  
 Year: 2022  
 Constituent Group: Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered: 19/08/2022

Date of Report Published: 01/09/2022

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Year Value	SDI	SI
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robonik ( Semi automatic / Fully Automatic )	1841	175.54	5.13	9.00	184 mg/dL	0.44	0.43
2	UREA II	Others ( Urease Berthelot / Ind Dye ) II	Any Analyser (Automation / Semi Automation )	39	28.92	13.99	4.05	62 mg/dL	0.70	1.83
3	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robonik ( Semi automatic / Fully Automatic )	1832	1.70	0.77	0.12	2 mg/dL	2.02	0.01
4	CHOLESTROL II	CHOD-PAP II	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.04	111 mg/dL	0.80	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Reprint Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>