

VIEW LAB MONTHLY SUMMARY

| | | | |
|---|---------------------------------|-------------------------------------|-----------------------|
| Lab Name | Govt. Block PHC, Chavadipalayam | Details About Robust Analysis | |
| Lab No | 7612 | Detail About Monthly Summary | |
| Month | August | Detail about SDI | |
| Year | 2023 | | |
| Constituent Group | Chemistry III | | |
| Click on the analyte to view Graphical Data | | All Analyser Result | Print |
| Date of Result Entered : 19/08/2023 | | | |
| Date of Report Published : 01/09/2023 | | | |

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 152 mg/dL | -2.62 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.9 mg/dL | 1.18 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 116 mg/dL | 0.47 | 0.22 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |

வட்டார மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 சாவடிப்பாளையம்
 திருப்பூர் மாவட்டம்-638 701