

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Near Ariyur Railway Gate, Thorapadi Details About Robust Analysis  
 Lab No 9297 Detail About Monthly Summary  
 Month July Detail about SDI  
 Year 2023  
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result Print


Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Sl. No.	Analyte	Method / Principle	Instrument	Mean	SD	CV	CV <sub>10</sub>	CV <sub>15</sub>	CV <sub>20</sub>	CV <sub>30</sub>
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.56 mg/dL	-0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	121 mg/dL	1.72	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

  
 Medical Officer  
 Urban Primary Health Center,  
 Thorapadi,  
 Vellore City Municipal Corporation