

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Near Ariyur Railway Gate, Thorapadi Details About Robust Analysis
 Lab No 9297 Detail About Monthly Summary
 Month May Detail about SDI
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/05/2023


Date of Report Published : 05/06/2023

Sl. No	Analyte	Principle	Equip. Make	Lot No	Particulars	CV	SP	Value	SDI	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	198 mg/dL	1.82	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	109 mg/dL	-0.96	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Home	View Lab Result	View Lab Monthly Summary
View Lab Result	Home	View Lab Monthly Summary


Medical Officer
 Urban Primary Health Center,
 Thorapadi,
 Vellore City Municipal Corporation