

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PG-1024

11:44 PM
Lab Name
Constituent Group
PT Item

Govt. PHC, Pudhuvalasai

Lab No

9261

Chemistry III

Date of Result Entered :

19/08/2023

Lyophilized human serum based

Date of Report Published :

01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	167.9 mg/dL	-0.85	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115.6 mg/dL	0.39	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Pudhuvalasai

THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, PUDHUVALASAI, RAMANATHAPURAM TALUK, RAMANATHAPURAM

Pudhuvalasai

TAMILNADU623522

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****