

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Near Ariyur Railway Gate, Thorapadi Details About Robust Analysis
 Lab No 9297 Detail About Monthly Summary
 Month April Detail about SDI
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/04/2023


Date of Report Published : 06/05/2023

Sl. No.	Analyte	Method / Principle	Manufacturer	Unit	Value	Reference Range	Delta	Delta %	Delta Z	Delta I
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	mg/dL	1832	119.45 - 6.04 - 7.21	122	0.35	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	mg/dL	1820	2.16 - 8.09 - 0.18	2.3	0.80	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	mg/dL	1750	111.61 - 5.19 - 5.79	110	-0.28	0.28	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Report Card	Lab B. Results	Chemical Results	Reporting Laboratory
Request Form	Request Form	Home	Sign-off


Medical Officer
Urban Primary Health Center,
Thorapadi,
Vellore City Municipal Corporation