

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :		CERTIFICATE NO	FICS/2023/PAND/VNR/09
UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE		EQUIPMENT NAME	CENTRIFUGE
PANDALKUDI,VIRUDHUNAGAR-626113		MANUFACTURE	LABOTECH
		MODEL	BDI-153
CALIBRATION DATE	15/02/2023	SERIAL NO	76600
CALIBRATION DUE DATE	14/02/2024	INSTALLED DEPARTMENT	LAB

1. General maintenance required Yes No
2. Any Physical damage observed Yes No
3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875

TEST REPORT

 Temperature 25°C
Humidity RH 55%



PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 1	1855	± 5	PASS
		Speed 4	2585	± 5	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :

Authorized Sign With Seal


*Electrical safety test result attached