

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Near Ariyur Railway Gate, Thorapadi Details About Robust Analysis  
 Lab No 9297 Detail About Monthly Summary  
 Month March Detail about SDI  
 Year 2023  
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl. No.	Analyte	Method - Principle Name	Apparatus / Instrument	Count	Mean	SD	CV	Result	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	98.81 mg/dL	1.07	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.32 mg/dL	-0.08	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	105 mg/dL	-1.02	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links



**Medical Officer**  
 Urban Primary Health Center,  
 Thorapadi,  
 Vellore City Municipal Corporation