

CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE PANDALKUDI, VIRUDHUNAGAR-626113		CERTIFICATE NO		TICS/2023/PAND/VNR/06	
		EQUIPMENT NAME		PIPETTE	
		MANUFACTURE MODEL		THERMOSCIENTIFIC	
		SERIAL NO		100-1000 µL	
CALIBRATION DATE	15/02/2023	INSTALLED DEPARTMENT		LAB	
CALIBRATION DUE DATE	14/02/2024				
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVC5PL22/07/884-01	27-JUL-2023		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.3	± 1	PASS
		1000	1000.2	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					