



CALIBRATION / TESTING REPORT

<b>CUSTOMER DETAILS :</b> GOVERNMENT PRIMARY HEALTH CENTRE PANDALKUDI, VIRUDHUNAGAR, TAMILNADU-626113		<b>CERTIFICATE NO</b>	TICS/2023/PAND/VNR/05
		<b>EQUIPMENT NAME</b>	MICROPIPETTE(10-100 µL)
		<b>MANUFACTURE MODEL</b>	THERMOSCIENTIFIC NA
<b>CALIBRATION DATE</b>	15/02/2023	<b>SERIAL NO</b>	QW16537
<b>CALIBRATION DUE DATE</b>	14/02/2024	<b>INSTALLED DEPARTMENT</b>	LAB

8.4

- 1. Have more dust in machine outside  Yes  No
- 2. Any Physical damaged in machine  Yes  No
- 3. Machine inside Cleaning (if required)  Yes  No

<b>TEST EQUIPMENT</b>	<b>MANUFACTURE</b>	<b>CERTIFICATE NUMBER</b>	<b>VALIDITY</b>
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPL22/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C  
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.1	± 1	PASS
		99	99.3	± 1	PASS

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE     
  CORRECTIVE MAINTENANCE REQUIRED     
  REMOVED FROM USE  
 Authorized Sign With Seal

Calibrated By :

