

CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS : UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE PANDALKUDI, VIRUDHUNAGAR-626113		CERTIFICATE NO	TICS/2023/PAND/VNR/07
		EQUIPMENT NAME	PIPETTE
		MANUFACTURE	THERMOSCIENTIFIC
		MODEL	100-1000 μ L
CALIBRATION DATE	15/02/2023	SERIAL NO	RW02952
CALIBRATION DUE DATE	14/02/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP22/07/884-01	27-JUL-2023

Temperature 25°C
Humidity RH 55%

TEST REPORT					
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	μ L	100	100.1	± 1	PASS
		1000	1000.1	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Authorized Sign With Seal

Calibrated By :




