

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kulakkaral Street	<a href="#">Details About Robust Analysis</a>	
Lab No	12258	<a href="#">Detail About Monthly Summary</a>	
Month	June	<a href="#">Detail about SDI</a>	
Year	2023		
Constituent Group	Chemistry III		

Click on the analyte to view Graphical Data

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95	290 mg/dL	0.19	3.21
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	38 mg/dL	0.80	1.42
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	2.7 mg/dL	0.73	0.04
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	116 mg/dL	0.30	0.76

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*D. P. S.*  
**MEDICAL OFFICER**  
**GOVT UPHC - IV**  
 Kulakkaral Street, Jamin Rayapet,  
 Pallavaram Municipality