

VIEW LAB MONTHLY SUMMARY

Lab Name: **Cort. Block PHC, Alangudi**

Lab No: **7263**

Month: **August**

Year: **2023**

Constituent Group: **Chemistry III**

[Details About Product Analysis](#)

[Details About Monthly Summary](#)

[Detail about SEH](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 17/08/2023

Date of Report Published : 01/09/2023

[Click on the analyte to view Graphical Data](#)

Sl.No	Analyte	Method / Reference Range	Analyser Name	No of Participants	DV	Participants		Test Value	SD	CV
1	GLUCOSE (R)	GDH (HLD) (R)	Holomak (Semi automatic / Fully Automatic)	1043	175.58	5.13	9.00	101.3 mg/dL	0.86	0.84
2	UREA (R)	UREASE (V) / GLDH (R)	Robomak (Semi automatic / Fully Automatic)	174	28.29	16.20	4.30	33 mg/dL	1.10	0.65
3	CREATININE (R)	Jaffe's Kinetic / Abalone Kinetic (R)	Robomak (Semi automatic / Fully Automatic)	1032	1.26	6.17	0.12	1.9 mg/dL	1.18	0.01
4	CHOLESTEROL (R)	CHOD-PAP (R)	Robomak (Semi automatic / Fully Automatic)	1764	133.80	4.07	4.04	110 mg/dL	0.82	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution Warning Signal						
Beyond ±3.0				Unacceptable performance Action Signal						

Links			
Home Address	Home Address	Lab Records	View Lab Monthly Summary
Home Summary	Change Password	Home	Support

Lab Name: Govt. Block PHC, Alangudi
 Lab No: 7263
 Month: May
 Year: 2023
 Constituent Group: Chemistry III

[Details About RoboNet Analyser](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE (B)	GDH POU (B)	Robonik (Semi automatic / Fully Automatic)	1054	100.00	5.19	6.39	170.2 mg/dl	-0.20	0.44
2	UREA (B)	Urease UV / GDH (B)	Robonik (Semi automatic / Fully Automatic)	163	27.01	13.39	3.74	33 mg/dl	1.36	0.59
3	UREA (TININ) (B)	Jaffes Kinetic / Alkaline Picrate (B)	Robonik (Semi automatic / Fully Automatic)	1054	1.76	7.48	0.13	1.73 mg/dl	-0.23	0.01
4	CHOLESTEROL (B)	CHOD PAP (B)	Robonik (Semi automatic / Fully Automatic)	1705	114.60	5.08	6.83	114 mg/dl	-0.10	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution / Warning Signal
Beyond ±3.0	Unacceptable performance / Action Signal

Name Govt. Block PHC, Alangudi
No 7263
Month June
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
					CV	SD		
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1523	286.63	5.17	14.02	268.3 mg/dL	-1.24
UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	30 mg/dL	0.63
CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	108.5 mg/dL	-1.03

SDI Range	Interpretation
within -1.00 to +1.00	Excellent
within ±1.01 to ±2.00	Good
within ±2.01 to ±2.99	Accept with caution. Warning Signal
beyond ±3.0	Unacceptable performance. Action Signal



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Alangudi
 Lab No: 2263
 Month: March
 Year: 2023
 Consultant Group: Chemistry II

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)
 Date of Result Entered : 20/03/2023
 Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE (F) B	GLUCOPOD B	Ribocnik (Semi automatic / Fully Automatic)	1678	91.45	7.46	6.52	96 mg/dL	0.66
2	UREA (F) B	UREAM (F) (2) (F) B	Ribocnik (Semi automatic / Fully Automatic)	175	48.20	11.70	6.46	31 mg/dL	-3.75
3	UREA (F) (F) B	Jafin Kinetic - Alkaline Picrate B	Ribocnik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.5 mg/dL	1.40
4	UREA (F) (F) B	UREA-PAP B	Ribocnik (Semi automatic / Fully Automatic)	1781	110.93	5.23	5.80	107 mg/dL	-0.68

CV Range	Interpretation
Within ± 0.0 to ± 1.00	Excellent
Within ± 1.01 to ± 2.00	Good
Within ± 2.01 to ± 4.00	Accept with caution - Warning Signal
Beyond ± 4.0	Unacceptable performance - Action Signal

Govt. Block PHC, Alangudi

7263

April

2023

Group Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
COSE III	GOD POD III	Robonik (Semi automatic / Fully Automatic)	1832	19.45	6.04	7.21	112.63 mg/dL	-0.95	0.34
A III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	52.75 mg/dL	2.31	0.70
CATININE III	Jaffe's Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.19 mg/dL	0.17	0.01
OLE STROL III	CHOD PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	125.57 mg/dL	2.41	0.28

SDI Range	Interpretation
0.00 to +1.00	Excellent
1.01 to ±2.00	Good
2.01 to ±2.99	Accept with caution. Warning Signal
±3.0	Unacceptable performance. Action Signal

NEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Atangudi	Details About Robust Analysis
Lab No	7263	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	113 mg/dL	0.60	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	52.9 mg/dL	0.26	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.22 mg/dL	1.11	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	113.8 mg/dL	0.17	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out