

CALIBRATION CERTIFICATE

Page : 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, PALLIYADI-629 169.	Certificate No:	VCPL/CC/23-24/28-001
	Date of Calibration:	29-09-2023
	Date of issue:	30-09-2023
	Date of Receipt:	29-09-2023
	Next cal Date:	28-09-2024

Details of Device Under Calibration

Description	CENTRIFUGE	Make:	B.D INSTRUMENTATION
Range	3500 RPM	Model/Type:	BDI-152
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	GPHC/LAB/CEN/01	Instrument Location:	LABORATORY

Environmental Conditions

Temperature	25.4 °C	Reference Standard No	SANAS TR4501
Humidity	53 % Rh	Calibration Procedure No	-

Method Used For Calibration

Details of Reference Standard Equipment For Calibration

S.No	Description	ID No / SL No	Certificate No	Validity
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005	SYSTEMS/ HTM 560 4-Dec-23

Calibration Result


S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	500	499.9	0.1	12
2	1000	999.6	0.4	12
3	1800	1802	-2	12
4	3000	2996	4	12
5	3500	3502	-2	12

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


R.GOPALAKRISHNAN
(CALIBRATION ENGINEER)



*** End of the Report ***

Authorised by:


M. BALAJI,
Quality Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyar Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Call: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com