



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Thelungankudikadu
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 175522
 Date of Result Entered: 20/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	CV	SD	Mean Value	SD
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.18	164 mg/dL	4.29
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	0.24
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.87	4.64	103 mg/dL	3.04

SDI Range	Interpretation
With in -1.00 to +1.00	Excellent
With in ±1.01 to ±2.00	Good
With in ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Thelungankudikadu
 Thelungankudikadu
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore


Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Thelungankudikadu
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No 12663
Date of Result Entered : 23/07/2023
Date of Report Published : 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.23
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.4 mg/dL	-1.23	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	111 mg/dL	-0.43	0.21

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Thelungankudikadu
Thelungankudikadu
TAMILNADU

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Thelungankudikadu
 Lab No: 12063
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analysar Name	No of Participants	DV	CV	Participants ML	Your Value	BDI	U
1	GLUCOSE III	GOD-POD III	Robonk (Beim automatic / Fully Automatic)	1023	280.03	5.17	14.12	251 mg/dL	-2.40	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate II	Robonk (Beim automatic / Fully Automatic)	1048	2.57	0.04	0.17	2.48 mg/dL	-0.53	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Beim automatic / Fully Automatic)	1004	113.73	4.46	5.07	103 mg/dL	-2.12	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution, Warning Signal.
Beyond +3.0	Unacceptable performance, Action Signal.

Links



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. PHC, Thelungankudikadu

Lab No

12663

Constituent Group Chemistry III

Date of Result Entered :

19/05/2023

PT item Lyophilized human serum based

Date of Report Published :

05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	188 mg/dL	0.76	0.41
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	1.62 mg/dL	-1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	117 mg/dL	0.41	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Thelungankudikadu
Thelungankudikadu
TAMILNADU

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102


Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. PHC, Thelungankudikadu
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No 12663
Date of Result Entered : 23/04/2023
Date of Report Published : 03/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	Participants		Your Value	SDA	U
					CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	6.04	7.21	127 mg/dL	1.05	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	8.09	0.18	2 mg/dL	-0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	5.19	5.79	115 mg/dL	0.59	0.23

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Thelungankudikadu
Thelungankudikadu
TAMILNADU

Pavani Christudoss
Dr. Pavani Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. PHC, Thelungankudikadu

Lab No

12663

Constituent Group Chemistry III

Date of Result Entered :

19/03/2023

PT Item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl. No.	Test Name	Method	Manufacturer	Lot No.	Expiry	Temp.	Result	Unit	Delta	Delta %
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1766	151.83	7.28	11.06	165 mg/dL	1.19	0.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1708	1.20	11.54	0.14	1.2 mg/dL	0.00	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1655	115.16	6.70	7.71	112 mg/dL	-0.41	0.38

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
thin ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Thelungankudikadu
Thelungankudikadu
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

P. Chikri
மருத்துவ அலுவலர்,
சுக அமைப்பு சுகாதார நிலையம்
தெலுங்குடிகுடி - 614623