



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 20.09.2023
		Calibration valid : 19.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE , PULIYUR, THIRUVANNAMALAI-635703.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINEPITTE – F3
iii	Value	100µl-1000µl
iv	Serial No	PHC/PU/PIP/002

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.04	500.02	1000.02
2	100.01	500.08	1000.10
3	100.02	500.09	1000.09
4	100.09	500.07	1000.12
5	100.05	500.05	1000.06
6	100.02	500.05	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)

