

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE, PALLIYADI-629 169	Certificate No:	VCPL/CC/23-24/28-003
	Date of Calibration:	29-09-2023
	Date of issue:	30-09-2023
	Date of Receipt:	29-09-2023
	Next cal Date:	28-09-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	THERMOSCIENTIFIC
Range:	( 10 to 100 ) $\mu$ l	Model/Type:	FINN PIPETTEF3
Resolution:	0.2 $\mu$ l	Condition of DUC	GOOD
Serial Number	RW01287	Location:	LABORATORY
ID number	-	Cal. At:	LAB
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	23.2 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	49 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION


S.NO	Duc Readings ( $\mu$ l )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	10.0	9.88	0.12
2	100.0	99.87	0.13
3	100.0	99.83	0.17

**Remarks:**

1. DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.

Calibrated By:

**R.GOPALAKRISHNAN**  
(CALIBRATION ENGINEER)



\*\*\* End of the Report \*\*\*

Authorised by:

**M. BALAJI,**  
Quality Manager

Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com