



# SILICON TECHNOLOGIES

2 & 3, Devi Illam, Priya Nagar,  
Telungupalayam Pirivu,  
Coimbatore - 641 039, TamilNadu.

GSTIN : 33APLPA1828A1ZM

T. ARUN PRASAD, M.Tech., MBA., (Ph.D)  
Managing Director  
Consultant Biomedical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment: **CENTRIFUGE** Protect Class/Type : Risk class I  
Make : KEMI S.No : 05090  
Model : C8C Location : **KATTABETTU - GPHC**

S.no	Measurement Parameters	Status	Remarks
I	<b>Electrical Test</b> (a). Safety Standards (range: 210-240 V) (b). Equipotential ground & Earthing (c). Risk of shock/fire Hazards, if any (d). Chassis & Insulation Resistance (e). Earth & Chassis Leakage Current (f). Battery condition & charging status	230volts 0 to 5volts No Zero ohms 10mA NA	Stabilizer/UPS  With in ranges well
II	<b>Physical Condition Test</b> (a). Cabinet & Visible damages, if any (b). Front panel/top lid, buckets, motor assembly & Power cord (c). Risk of mechanical damages, if any	No Ok  No	Checked ok
III	<b>Performance Test</b> (a). On/off switch (b). Speed regulator switch & functional tests (c). Running test (d). Abnormalities (e). Display checkup (f). Keypad/indicators & connectors	Normal ok accurate Ok No Well Well	ok perfect perfect
IV	<b>Calibration Test</b> (a). Vibration test (b). Speed regulation gradual hike/wobbling tests (c). Running test with load (d). Display checkup (f). timer test	Done ok ---	perfect
<b>Formulated By</b> : Einstein Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>
<b>Endorsed By</b> : Research team head		 <b>SILICON TECHNOLOGIES</b> A Biomedical Instrumentation Company Coimbatore - 641 039 T. Arunprasad Cell: 98652 85754	
Reviewed On : <b>07.03.23</b>			
Next Review : <b>07.09.23</b>			