

VIEW LAB MONTHLY SUMMARY


Lab Name	Govt. UPHC, Kulakkarai Street	Details About Robust Analysis	
Lab No	12258	Detail About Monthly Summary	
Month	January	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		
Click on the analyte to view Graphical Data		All Analyser Result	Print
Date of Result Entered : 19/01/2023			
Date of Report Published : 09/02/2023			

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	73	122.26	6.20	7.58	131 mg/dL	1.15	1.77
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	53	42.30	10.70	4.53	37 mg/dL	-1.17	1.24
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	2.23	9.52	0.21	2.3 mg/dL	0.33	0.05
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	111.59	5.08	5.67	107 mg/dL	-0.81	0.93

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


MEDICAL OFFICER
GOVT. UPHC - IV
 Kulakkarai Street, Jamin Rayapat,
 Pallevaram Municipality