



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Kulakkarai Street
 Lab No: 12238
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Results](#)

[Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	3.46	9.31	181 mg/dL	9.76	2.38
3	UREA III	Others (Ureaase Barbitolot / Nod Dye) III	Any Analyser (Automation / Semi Automation)	59	28.92	13.99	4.85	30 mg/dL	8.27	1.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.78	7.18	8.12	1.7 mg/dL	-8.48	8.83
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.96	4.36	4.97	113 mg/dL	8.21	8.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Quality Assurance	Quality Methods	Submit Lab Results	View Lab Monthly Summary
Quality Summary	Change Password	Home	Sign-out

Best Viewed at pixels 1024 X 768

MEDICAL OFFICER
GOVT UPHC - IV
 Kulakkarai Street, Jamin Rayapet,
 Pallavaram Municipality