

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Thelungankudikadu
 Lab No: 12063
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analysar Name	No of Participants	DV	CV	Participants ML	Your Value	BDI	U
1	GLUCOSE III	GOD-POD III	Robonk (Beini automatic / Fully Automatic)	1023	280.03	5.17	14.12	251 mg/dL	-2.40	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate II	Robonk (Beini automatic / Fully Automatic)	1048	2.57	0.04	0.17	2.48 mg/dL	-0.53	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Beini automatic / Fully Automatic)	1004	113.73	4.46	5.07	103 mg/dL	-2.12	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution, Warning Signal.
Beyond +3.0	Unacceptable performance, Action Signal.

Links