

April - 2023

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Arattavadal	Details About Robust Analysis								
Lab No	12912	Detail About Monthly Summary								
Month	April	Detail about SDI								
Year	2023									
Group	Chemistry III									
Click on the analyte to view Graphical Data										
All Analyser Result Print										
Date of Report Published : 19/04/2023										
Date of Report Published : 06/05/2023										
SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
	GLUCOSE III		Robonik (Semi Automatic)	112	101.47	0.14	0.21	120.5 mg/dL	0.17	0.00
	CREATININE III	Alkaline Picrate III	Robonik / Semi Automatic / Fully Automatic	1050	2.10	0.00	0.10	1.8 mg/dL	2.00	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik / Semi Automatic / Fully Automatic	1750	111.61	5.19	5.79	116.2 mg/dL	0.79	0.28
SDI Range		Interpretation								
Within ± 1.00 to ± 1.99		Excellent.								
Within ± 2.01 to ± 2.99		Good.								
Within ± 3.00 to ± 3.99		Accept with caution. Warning Signal.								
Beyond ± 4.00		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out