

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Kulakkarai Street

Details About Robust Analysis

Lab No 12258

Detail About Monthly Summary

Month February

Detail about SDI

Year 2023

Constituent Chemistry III

Group

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 22/02/2023

Date of Report Published : 08/03/2023

| Sl.No | Analyte        | Method / Principle Name                   | Analyzer Name                               | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|-------------------------------------------|---------------------------------------------|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                           |                                             |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                               | Any Analyser (Automation / Semi Automation) | 67                 | 156.09 | 5.17         | 8.07 | 152 mg/dL  | -0.51 | 1.97 |
| 2     | UREA III       | Others ( Urease Berthelot / Ned Dye ) III | Any Analyser (Automation / Semi Automation) | 46                 | 32.73  | 12.34        | 4.04 | 29 mg/dL   | -0.92 | 1.19 |
| 3     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III     | Any Analyser (Automation / Semi Automation) | 70                 | 3.39   | 9.61         | 0.33 | 3.8 mg/dL  | 1.26  | 0.08 |
| 4     | CHOLESTROL III | CHOD-PAP III                              | Any Analyser (Automation / Semi Automation) | 143                | 112.46 | 5.14         | 5.78 | 122 mg/dL  | 1.65  | 0.97 |

| SDI Range             | Interpretation                           |
|-----------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

Links

|                |                 |                    |                          |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods  | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home               | Sign-out                 |

**P**  
**MEDICAL OFFICER**  
**GOVT UPHC - IV**  
 Kulakkarai Street, Jamin Rayapet,  
 Pallavaram Municipality