

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Kulakkarai Street

Lab No 12258

Month April

Year 2023

Constituent Group Chemistry / III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Click on the analyte to view Graphical Data

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83	123 mg/dL	0.68	1.59
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation)	56	43.27	11.10	4.80	43 mg/dL	0.06	1.28
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15	2.2 mg/dL	0.40	0.04
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	112 mg/dL	0.16	0.80
Interpretation										
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

**LAB OFFICER**  
**MEDICAL OFFICER**  
 GOVT UPHC - IV  
 Kulakkarai Street, Kullukkarai  
 Palayamkottam, Tiruppur