

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kulakkarai Street	Details About Robust Analysis
Lab No	12258	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/03/2023


Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	92.52	7.71 7.13	92 mg/dL	-0.07	1.86
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	50.51	14.61 7.38	56 mg/dL	0.74	2.09
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	70	1.32	7.52 0.10	1.4 mg/dL	0.81	0.02
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29 4.78	12 mg/dL	0.09	0.79

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out


MEDICAL OFFICER
GOVT UPHC - IV
Kulakkarai Street, Jamin Rayapat,
Pallavaram Municipality