



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - AUGUST 2023



MS 1004

Lab Name: Govt. PHC, Thejangan-kudikadu  
 Consultant Group: Chemistry III  
 PT Item: Lyophilized human serum based

Lab No: 12083  
 Date of Result Entered: 20/08/2023  
 Date of Report Published: 01/09/2023

Item No.	Method	Participant Name	Accepted Name	No. of Participants	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	178.88	5.13	9.16	184 mg/dl
2	CREATININE III	Jaffes Kinetic - Alkaline Phorate III	Robonik ( Semi automatic / Fully Automatic )	1882	1.76	6.77	9.12	1.8 mg/dl
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.14	189 mg/dl

SDI Range	Interpretation
With in -1.00 to +1.00	Excellent.
With in ±1.01 to ±2.00	Good.
With in ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. PHC, Thejangan-kudikadu  
 Thejangan-kudikadu  
 TAMILNADU

Coordinator Contact Details:  
 Email: [clinc@cmcvellore.ac.in](mailto:clinc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Tamara Chidambaram*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential