


CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE PANDALKUDI,VIRUDHUNAGAR-626113			CERTIFICATE NO TICS/2023/PAND/VNR/08		
			EQUIPMENT NAME PIPETTE		
			MANUFACTURE MODEL DIAGNOPETTE IQMI		
CALIBRATION DATE 15/02/2023		SERIAL NO PAND-08			
CALIBRATION DUE DATE 14/02/2024		INSTALLED DEPARTMENT LAB			
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		MANUFACTURE		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		SHIMADZU		TVCSPL22/07/884-01	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.4	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 			Authorized Sign With Seal 