

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Veeralur	Details About Robust Analysis	
Lab No	12923	Detail About Monthly Summary	
Month	May	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		


Click on the analyte to view Graphical Data

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	176 mg/dL	-0.52	0.44
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	15	1.76	8.24	0.14	1.6 mg/dL	-1.10	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	124 mg/dL	1.61	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


 MEDICAL OFFICER
 Govt. Primary Health Centre
 Veeralur-606 901

Links