



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/11986	MEDICAL DEVICES	Page No:1 of 1
Issue Date	29-09-2023	
SRF No & Date	SRF/23/00453-0001 & 28-09-2023	
Receipt Date	28-09-2023	
Calibration Date	28-09-2023	
Calibration Due	27-09-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
ESWARACHETTIPALAYAM-638702, TIRUPPUR DISTRICT.

**Details of Device Under Calibration (DUC)**

Description	ELECTRICAL SAFETY(MICROSCOPE)	Make & Model	LABOMED & VISION 2000
Range	MULTI	Sr. No	220135397
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY**

**RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 87	Uncertainty in % ( ± ) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 195	Uncertainty in % ( ± ) 9.3
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 241	Uncertainty in % ( ± ) 8.4

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

*KSH*  
(Calibration Engineer)  
K.SATHYAMOORTHY



*C.S*  
Technical Manager  
(C.SHANMUGARAJ)

Chief Executive