



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name	Govt. Block PHC, Sayalkudi	Lab No	4594
Constituent Group	Chemistry III	Date of Result Entered :	20/03/2023
PT item	Lyophilized human serum based	Date of Report Published :	04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	95.87 mg/dL	0.64	0.31
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	50.51	14.61	7.38	61.66 mg/dL	1.51	2.09
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.33 mg/dL	0.00	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	113.09 mg/dL	0.37	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Sayalkudi
GOVERNMENT PRIMARY HEALTH CENTRE , RAMANATHAPURAM.D.T,
Sayalkudi
TAMILNADU623120

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

V. L. Karthikeyan

மருத்துவ அலுவலர்
மெட்ரிக் கட்டுப்பாட்டு அமைச்சு, தமிழ்நாடு
பெங்களூர்-623 120.
இராமநாதபுரம் மாவட்டம்.