



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2022



PC-1024

Lab Name                      GOVT.PRIMARY HEALTH CENTRE                      Lab No    7518  
 Constituent Group              Chemistry III                                      Date of Result Entered :                      20/04/2022  
 PT Item                              Lyophilized Serum                                      Date of Report Published :                      05/05/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1781	263.81	7.17	18.93	225 mg/dL	-2.05	0.90
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1758	4.67	9.43	0.44	8.7 mg/dL	9.16	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1741	115.66	6.19	7.16	123 mg/dL	1.03	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 GOVT.PRIMARY HEALTH CENTRE  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.  
 VETTAVALAM  
 TAMILNADU606754

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*ky*  
 Medical Officer  
 Govt.Primary Health Centre  
 Vettavalam - 606 754  
 Tiruvannamalai District