

VIEW LAB MONTHLY SUMMARY

Lab Name GOVT.PRIMARY HEALTH CENTRE
 Lab No 7518
 Month June
 Year 2022
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print


Date of Result Entered : 20/06/2022

Date of Report Published : 07/07/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1828	87.43	7.74	6.77	84 mg/dL	-0.51	0.32
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1769	1.99	8.26	0.16	1.8 mg/dL	-1.16	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1737	118.45	6.70	7.93	112 mg/dL	-0.81	0.38

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out 

Medical Officer
 Govt.Primary Health Centre
 Vettavalam - 606 754
 Tiruvannamalai District