



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2022



Lab Name Govt.UGPHC, Vettavalam Lab No 7518
Constituent Group Chemistry III Date of Result Entered : 20/08/2022
PT item Lyophilized Serum Date of Report Published : 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1833	104.52	7.14	7.46	102 mg/dL	-0.34	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1780	6.46	8.21	0.53	6.3 mg/dL	-0.30	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1717	115.04	5.56	6.39	112 mg/dL	-0.48	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt.UGPHC, Vettavalam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.
Vettavalam
TAMILNADU606754

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Joe
Medical Officer
Govt. Primary Health Centre
Vettavalam - 606 754
Tiruvannamalai District