

	CAL	LIBRAT	ΓΙΟ	N CER	TIFIC	CA	TE_		
CERTIFICATE NO: SBS/CL/23/12158			MEDICAL DEVICES P					Page	No:1 of 1
Issue Date			30-0	9-2023				90	110.1 01 1
SRF No & Date			SRF	/23/00494-00	01 & 29	9-09-	2023		
Receipt Date			100000000000000000000000000000000000000	9-2023					
Calibration Date			29-0	9-2023					
Calibration Due			28-09-2024						
Customer Name & Address									
GOVERNMENT PRIMARY H	EALTH CENTR	RE,							
SAYALKUDI-623120, RAMANA	THAPURAM (DISTRICT.							
		Details of D	evice	Under Calibra	ation (DU	C)			
Description : SEMI AUTO ANALYZER		Make & Model : ROBONIK & PRIE			ROBONIK & PRIETES	TEST TOUCH			
Range : MULTI		Sr. No :			ATCD3511220RBK				
esolution : MULTI		Identification No			NA				
UC Condition : SATISFACTORY		Location		:		LABORATATORY			
	Enviro	nmental Cond	litions	& Calibration	Proced	ure D	etails		
Environmental Details Temperature:25.6 ° C		Relative Humidity		52% RH					
Calibration Procedure No SBS/CP/MD/20		Calibration done at		ONSITE					
		Refer	ence	Standards De	tails				
S.No Description	S.No Description Make/ SI No:				Certificate No				Validity
1 Electrical Safety Analyser Rigel Medica		l & 44L-1059		M-230809-16-4				10-08-2024	

	ELECTRICAL SAFETY							
RESULTS CONTROL OF THE PROPERTY OF THE PROPERT								
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)					
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)					
	>20 M Ω	97	13.92					
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)					
	<5000µAfor B,BF,CF	218	7.2					
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)					
	<500µAfor B,BF,CF	235	7.2					

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards. Calibrated By

(Calibration Engineer)
M.DINESH





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS