

CERTIFICATE NO	CATE NO: SBS/CL/23/12159			MECHANICAL		Pa	age No:1 of 1	
Issue Date				30-09-2023				
SRF No & Date				SRF/23/00494-0002 & 29-09-2023				
Receipt Date				29-09-2023				
Calibration Date				29-09-2023				
Calibration Due				28-09-2024				
Customer Name &								
GOVERNMENT P	RIMARY HEALT	H CENTRE,						
SAYALKUDI-62312	20,RAMANATHA	PURAM DIST	RICT.					
		Det	ails of D	evice Under C	alibration (OUC)		
Description : CENTRIFUGE			Make & Model :		MICROSIL & MK-100			
Range	: 500	0 RPM		Sr. No		:	276/386	
Resolution : 100 RPM				Identification No :		NA		
DUC Condition : Satisfactory				Location :		LABORATORY		
	En	vironmental C	Condition	ns & Standard	Operating F	rocedu	ire Details	
Environmental Details Temperature: 24.3			e: 24.3 °	°C Relative Humidity		54 % Rh		
Calibration Procedure No SBS/CP/ML/04			/04	Calibration done at		ONSITE		
			Refer	ence Standard	s Details			
S.No Description	Description Make			/ SI No: C		Certificate No		Validity
Digital Tachometer LINE S			LINE SE	EIKI / 175-0034	JRPI	JRPM-CCTR-A&S-2023-0013		09-06-2024
CALIBRATION RE	SULTS		1					

DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
RPM	RPM	RPM	%	
1000	999.8	0.2	5.78	
2000	1999.9	0.1	5.78	
3000	2999.7	0.3	5.78	
	CALIBRATION READINGS RPM 1000 2000	CALIBRATION READINGSSTANDARD INSTRUMENTSRPMRPM1000999.820001999.9	CALIBRATION READINGSSTANDARD INSTRUMENTSDEVIATIONRPMRPMRPM1000999.80.220001999.90.1	

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) M.DINESH



Quality Manager (C.SIVABALAN)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

Authorised Signatory