

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered:

20/03/2023

PT item

Lyophilized human serum based

Date of Report Published:

04/04/2023

SLN	. Amelia	Method / Principle	A make a mark Manage	No of	DV	Participants		Your	ant	
31.IV	o Analyte	Name	Analyzer Name	Participants	DV	CV	SD	Value	SDI	U:
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	96 mg/dL	0.66	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	102 mg/dL	-1.54	0.28

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS :

Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******

MEDICAL OFFICER
PRIMARY HEALTH CENTER
MARUTHUR,
MARUTHUR,
MARIYALIBOD TO



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered :

20/04/2023

PT item

Lyophilized human serum based

Date of Report Published:

06/05/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	U
	A)	Name		rancipants		CV	SD	Value		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	126 mg/dL	0.91	0,34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	124 mg/dL	2.14	0.28

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS :

Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******

MEDICAL OFFICER
PRIMARY HEALTH CENTER
MARUTHUR,
MARUTHUR,
ARIYALUR(D)



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered:

20/05/2023

PT item

Lyophilized human serum based

Date of Report Published:

05/06/2023

SI.Ne	Acabasa	Method / Principle	A	No of	DV	Participants		Your	ani	-1/1
91.14	o Analyte	Name	Analyzer Name	Participants	DV	CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	139 mg/dL	-4.46	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0 Unacceptable performance. Action Signal.						

LAB ADDRESS:

Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered:

19/06/2023

PT item

Lyophilized human serum based

Date of Report Published:

04/07/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Partic	dipants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	301 mg/dL	0.97	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0 Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered:

20/07/2023

PT item

Lyophilized human serum based

Date of Report Published:

07/08/2023

SI.Ne	Analyte	Method / Principle	A mala man Name	No of	57.3.5	Participants		Your		
21.141	2 Analyte	Name	Analyzer Name	Participants	DV	CV	SD	Value	SDI	U
4	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	120 mg/dL	1.76	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0 Unacceptable performance. Action Signal.						

LAB ADDRESS : Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss **CMC EQAS Coordinator** Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report ******

ANDIMADAM(T.K), ARIYALUR(D.T)



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name

Govt. PHC, Maraudhur

Lab No

12005

Constituent Group

Chemistry III

Date of Result Entered:

19/08/2023

PT item

Lyophilized human serum based

Date of Report Published:

01/09/2023

SLN	o Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Partic	ipants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	170 mg/dL	-0.62	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS :

Govt. PHC, Maraudhur THE MEDICAL OFFICER, maruthur Marudhur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss
Dr. Pamela Christudoss

CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******

MEDICAL DEFICER

PRIMARY HEALTH CENTER

MARUTHUR,

MARUTHUR,

ANDIMADAM(T.K), ARIYALIPO