

				CAL	IBRAT	101	I CER	TIFIC	CA	TE	
CERTIFICATE NO: SBS/CL/23/12378						MEDICAL DEVICES					Page No:1 of 1
ssue Dale						30-09-2023					
SRF No & Date						SRF/23/00520-0003 & 30-09-2023					
Receipt Date						30-09-2023					
Calibration Date						30-09-2023					
Calibration Due						29-09-2024					
Custo	mer Name	dress			25-0	25-03-2024					
				RY HEALTH	CENTRE						
NERL	JPERICHA	L-64160	02,TIRUF	PUR DISTR	ICT.						
					Details of De	vice L	nder Calibra	tion (DI	JC)		
Description			ELECTRICAL SAFETY (MICROSCOPE)					,			
						Make & Model				LABOMED & VISION 20	
Range ;			MULTI			Sr. No			:	230146964	
DUC Condition			MULTI			Identification No			27	NA	
DUC (Condition	_:_	SATISE	ACTORY		Local			:	LABORATORY	
					mental Condi	tions	& Calibration	Proced	iure C	etalls	
Environmental Details Temperature:25.6*					25.6°C	5.6°C Relative Hur		nidity 52% RH		52% RH	
Calibration Procedure No SBS/CP/MD/2				29 Calibration of		one at ONSITE					
					Refere	nce S	tandards De	talls			
S.No	No Description Make/ SI					lo:		Certificate No			Validity
1	1 Electrical Safety Analyser R					Rigel Medical & 44L-1059			M-230809-16-4		

ELECTRICAL SAFETY RESULTS S.no **SPECIFICATION MEASURED VALUES EXPANDED UNCERTAINTY (±)** 1 Insulation Resisitance Measured values in MΩ Uncertainty in % (±) >20MΩ 88 13.92 2 Earth Leakage Measured values in µA Uncertainty in % (±) <5000µAfor B,BF,CF 119 8.1 Measured values in µA Uncertainty in % (±) 3 Enclosure Leakage <500µAfor B,BF,CF 236

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer)

M.RAGUL

Authorised Signatory

Technical Manager

C.SHANMUGARAJ