

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12379	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00520-0004 & 30-09-2023	
Receipt Date	30-09-2023	
Calibration Date	30-09-2023	
Calibration Due	29-09-2024	

Customer Name & Address
 GOVERNMENT URBAN PRIMARY HEALTH CENTRE,
 NERUPERICHAL-641602, TIRUPPUR DISTRICT.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY (CELL COUNTER)	Make & Model	SYSMEX & 100
Range	MULTI	Sr. No	B5850
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 97	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B,BF,CF	Measured values in µA 125	Uncertainty in % (±) 12.3
3	Enclosure Leakage <500µA for B,BF,CF	Measured values in µA 229	Uncertainty in % (±) 8.6

- REMARKS**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
 - Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


 (Calibration Engineer)
 M.RAGUL




 Technical Manager
 C.SHANMUGARAJ

Chief Executive

