

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Ilavanasurkottai

Lab No: 7294

Month: April

Year: 2023

Constituent: Chemistry III

Group: Chemistry III

Details About Robust Analysis

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic)	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Pirate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110 mg/dL	-0.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#) [Modify Methods](#) [Submit Lab Results](#) [View Lab Monthly Summary](#)

[Yearly Summary](#) [Change Password](#) [Home](#) [Sign-out](#)

MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 ELAVANASUR KOTTAI - 607202
 KALLAKURICHI DISTRICT