




RAJA MEDICAL SYSTEMS

No.15, V.O.C STREET, COIMBATORE - 641 011. MOBILE : 96263 13796.

Calibration Protocol & Report for Medical Equipment

Equipment : REFRIGERATOR
MAKE : LG
MODEL : INTELLOCOOL

PROTECT CLASS /TYPE : Class II
S. No. : G1248LAG
Location : GPHC, KAPPALA

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current	230 volts 0 to 5 volts No Zero ohms 10mA	Stabilizer/UPS With in Ranges
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (a) Regular functional mode / cooling temp (b) Key pad, display and settings (c) Abnormalities (d) Knobs & closures alignment	Normal OK No Ok	Well Well
4	Calibration Test (a) Cooling mode test (b) Normal mode test (c) Temp test	Checked Done	tested
Formulated By: Einstein Research Laboratory		Issued By:	User Department:
Endorsed By : Research team head 		RAJA MEDICAL SYSTEMS No.15, V.O.C STREET, COIMBATORE - 641 011.	
Reviewed On: 10.02.23 Next Review : 09.02.24			