

				CALI	BRA	TION	CERT	IFICAT	<u>E</u>	
CERTIFICATE NO: SBS/CL/23/12062					ME	MECHANICAL		Page No:1 of 1		
Issue Date						30-09-2023				
SRF No & Date					SRF/23/00469-0002 & 29-09-2023					
Receipt Date						29-09-2023				
Calibration Date						29-09-2023				
Calibration Due						28-09-2024				
Custome	er Name	& Add	ress							
GOVERN	NMENT P	RIMAR	RY HEALTH C	ENTRE,						
NATHAN	I,RAMAN	ATHA	PURAM DISTR	RICT.						
				Det	ails of D	evice Un	der Calibrat	tion (DUC)		
Descripti	ion	5	CENTRIFUC	3E		Make &	Model	;	SOLO KRAFT 8	NA
Range :		3500 RPM		Sr. No			AG29E			
Resolution :		1 R	1 RPM		Identification No			NA		
DUC Condition : Satisfactory			Location			:	LABORATORY			
			Enviro	onmental C	ondition	ns & Star	ndard Opera	iting Procedu	re Details	
Environmental Details Temperature: 24					2: 24.3 %	24.3 °C Relative Hu		midity 54 % Rh		
Calibration Procedure No SBS/CP/ML/04					/04	4 Calibration		done at ONSITE		
					Refer	rence Sta	ndards Det	ails		
S.No D	o Description Ma			Make/ S	Make/ SI No:		Certificate No		Validity	
1 Di	Digital Tachometer				Lutron & DT-2238		JRPM-CCTR-A&S-2022-0036		03-10-2023	

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.8	0.2	4.2	
2	2000	1999	1	4.2	
3	3000	2999	1	4.2	

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

(Calibrated By
(Calibration Engineer)
P.PRASANNA



Quality Manager (C.SIVABALAN)

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS